

# In-house Laboratory Quality Check List



	PRE-ANALYTICAL FACTOR	WHAT TO CHECK
1	SAMPLE QUALITY	Did the sample appear normal and free of interfering substances (e.g. haemolysis, lipaemia, clots) before the test was run?
2	SAMPLE HANDLING	Was the sample collected, stored and handled according to the analyser's instruction manual? Were the correct blood tubes used? Were the blood tubes filled correctly? Was the test run in the required time frame? Was the Haematology sample mixed well (i.e. inverted 10 times) immediately before testing?
3	ANALYSER OPERATION	Was the analyser operated according to its instruction manual? Were all consumables and accessories appropriate and within their expiry date?
4	ANALYSER MAINTENANCE	Has the analyser been well maintained and cleaned according to its instruction manual?
5	QUALITY CONTROL (QC)	Have all required QC tests been run and passed?

## REMEMBER - Confirmatory tests are available

Submit samples to your local SVS Pathology Network veterinary laboratory (fees apply):

- Vetpath Laboratory Services (WA, NT)
- ASAP Laboratory (Vic)
- Vetnostics (NSW)
- QML Vetnostics (QLD, northern NSW)
- TML Vetnostics (Tasmania)

## SVS - Specialist Veterinary Services Pathology Network



asaplab.com.au  
1300 838 522



vetqml.com.au  
1300 838 765



tmlpath.com.au  
1300 838 765



vetnostics.com.au  
02 9005 7468



vetpath.com.au  
08 9317 0777